Design and Holographic Field Reconstruction of Ultrasonic Lenses for Drug Delivery in non-Human Primates

Diana Andrés[1], **Alicia Carrión**[1], **Nathalie Lamothe**[1], **José A. Pineda-Pardo**[2], **Noé Jiménez**[1], **Francisco Camarena**[1]

[1]Instituto de Instrumentación para Imagen Molecular, CSIC - Universitat Politècnica de València, Spain

[2]HM CINAC, Fundación HM Hospitales de Madrid, University Hospital HM Puerta del Sur. Madrid, Spain


INTRODUCTION

Background

Transcranial focused ultrasound is emerging as a reliable and non-invasive therapeutical approach for many neurological disorders. However, in its way to the central nervous system, ultrasound beams experience strong aberrations due to the stiff and irregular cranial bones. Acoustic holograms have been used to correct these aberrations and shape the acoustic beam.

Our proposal

In this work we use **acoustic holographic lenses** coupled to a custom-made large aperture focused transducer (140-mm radius of curvature and 100-mm aperture) at 500 kHz for blood brain barrier opening in the left posterior putamen of rhesus macaques. We evaluate the system performance making direct measurements with an ex-vivo macaque skull. The focal region also was reconstructed by **holographic projection techniques** to avoid reflections between the skull and the hydrophone.



METHODS

Simulations


Pseudo-spectral (*k*-space) time-domain simulations of the constitutive relations for acoustic waves in heterogenous media including frequency-power law absorption and dispersion are used. The transducer was modelled as a quasi-rigid surface for a better approach to reality, and simulations ran during long time to capture multiple reflections in the skull and between the skull and the transducer.


Holographic lens design

The **holographic lens** was designed with time reversal methods, setting a virtual source at the target location. Resulting holographic lens is designed assuming each pixel of the lens vibrates as a Fabry-Pérot resonator, encoding the phase of the wave. The resulting lens is 3D printed with photo-sensitive polymers.

Experiments


Direct acoustic measurements were done with a piezoelectric hydrophone in a degassed water tank. **Holographic projection techniques** were also applied to retrieve the whole focus shape. To do this, an XY-plane located 8 mm above the focus is measured. The plane dimensions (30x30 mm^2) must ensure retrieving all the information coming from the transducer and the formed focus.







RESULTS

Acoustic fields obtained by simulation, direct measurement, and holographic projection were in good agreement. Focus sizes considering a -3 dB threshold were 3 x 3 x 14 mm^3 in the simulation, 4.5 x 4.5 x 15.5 mm^3 in direct measurements and 4 x 4 x 14 mm^3 using the holographic projection.



The focal spot appeared at the target location for simulation and both measurements, generating 0.5 MPa rarefaction pressure, with 55% and 40% of the focal volume lying in the left putamen region for simulation and holographic reconstruction, respectively. A 14% of the target volume was sonicated in the simulation results, while 38% of this volume was actually sonicated, measured using **holographic reconstruction**.





CONCLUSIONS

Acoustic holograms are a robust method to adapt large-focused transducer focus to a small target in the near field of the transducer, even when the focus needs to be steered. **Holographic projection** of the measured data match with the direct measured ones, which allows to reconstruct 3D volumes very accurately with a single measurement, avoiding artifacts that can be produced for the presence of the hydrophone near the skull.

REFERENCES

[1] Treeby, B. E., et al. *J. Acoust. Soc. Am*. 131(6), 4324-4336 (2012).

[2] Maimbourg, G., et al. Phys. Med. Biol., 63(2), 025026 (2018).

[3] Jiménez-Gambín, S., et al. *Phys. Rev. Appl*., 12(1), 014016 (2019).

[4] Sapozhnikov, Oleg A., et al. J. Acoust. Soc. Am., 138(3), 1515-1532 (2015).

We thank financial support from Spanish Administration through grants IJC2018-037897-I, FJC2019-040453-I and FPU19/00601, Generalitat Valenciana through the programs Prometeo CIPROM/2021/003, FDEGENT/2019/004, Project Emergentes CIGE/2021/175 and Equipamiento e Infraestructuras FEDER 2021-22 IDIFEDER/2021/004, and Agència Valencia de la Innovació INNVA2/2022/11.

Single element transducers | Phased-arrays systems | 3D Printed Lenses | Acoustic holograms

2022 IEEE IUS International Ultrasonics Symposium Venice, Italy || October 10-13, 2022

i3M Instituto de Instrumentación para Imagen Molecular

GENERALITAT VALENCIANA | TOTS A UNA veu

AVI AGÈNCIA VALENCIANA DE LA INNOVACIÓ